

Corrigendum to the Quotation Call Notice (Bid Reference No.BC/QUM/008/2024/403/dt.03.09.2024 relating to **SUPPLY & INSTALLATION OF CELL SEPARATOR (APHERESIS) MACHINE TO CENTRAL RED CROSS BLOOD CENTRE, CUTTACK**

<i>In the matter of</i>	<i>Sl No</i>	Present terms and conditions	Revised terms and conditions
<i>Regarding Technical specification as stated at Section – V</i>	3	Built in automated protocols for majority (4 of 6) of the below procedures. a. Leukoreduced Plasma Collection. b. Therapeutic Plasma Exchange. c. Single or double RBC collection and/or RBC Exchange. d. Peripheral Blood Stem Cell Collections. e. Granulocyte Collection. f. Leukoreduced platelet collection..	Leukoreduced platelet collection.
	5	Automated Self-test to ensure maximum patient Safety.	Automated Self-test to ensure maximum Donor Safety.
	9	Built in Bolus management function.	System should be able to automatically add PAS Solution to platelet product.
	10	Single access line for performing procedure of plateletpheresis. Double access line for performing procedures of therapeutics plasma exchange/therapeutic erythrocytapheresis/therapeutic leucopheresis.	System should have ECV less than 200 ml. in single needle.
	16	Should have Custom prime management (e.g. blood prime).	(OMIT)
	21	Online UPS of suitable rating with voltage regulation, spike protection and maintenance free batteries for 60 minutes back up to be provided.	Online UPS of suitable rating with voltage regulation, spike protection and maintenance free batteries for 20 minutes back up to be provided.

Other terms and conditions mentioned in the quotation call notice remain unaltered.

Jones 17.9.24
 Director, CRCBC, Cuttack

Copy Communicated / placed in website of CRCBC (www.crcbloodcentrectc.org) & Notice board of CRCBC, Cuttack

Jones 17.9.24
 Director, CRCBC, Cuttack
Director
Central Red Cross Blood Centre
Cuttack